



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Electrical Calibration (As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

December 11, 2020

Issue Date:

April 30, 2025

Expiration Date:

April 30, 2027

Accreditation No.:

110993

Certificate No.:

L25-333

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|--|---|--|--|---|----------------------|
| Electrical | Equipment to Measure DC Voltage | Up to 100 mV | $9 \mu\text{V/V} + 3 \mu\text{V}$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS02 | F, O |
| Electrical | Equipment to Measure DC Voltage | 100 mV to 1 V | $4 \mu\text{V/V} + 1 \mu\text{V}$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS02 | F, O |
| Electrical | Equipment to Measure DC Voltage | 1 V to 10 V | $4 \mu\text{V/V} + 1 \mu\text{V}$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS02 | F, O |
| Electrical | Equipment to Measure DC Voltage | 10 V to 100 V | $5 \mu\text{V/V} + 1 \mu\text{V}$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS02 | F, O |
| Electrical | Equipment to Measure DC Voltage | 100 V to 1 000 V | $6 \mu\text{V/V} + 1 \mu\text{V}$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS02 | F, O |
| Electrical | Equipment to Measure DC Voltage | 1 000 V to 10 000 V | 0.35 % of Reading | Fluke 8588 A DMM HP 3458A DMM Fluke 80 Divider | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | Up to 100 mV | $11 \mu\text{V/V} + 2 \mu\text{V}$ | Fluke 5700A Multi-Function Calibrator | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | 100 mV to 1 V | $10 \mu\text{V/V} + 3 \mu\text{V}$ | Fluke 5700A Multi-Function Calibrator | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | 1 V to 10 V | $9 \mu\text{V/V} + 7 \mu\text{V}$ | Fluke 5700A Multi-Function Calibrator | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | 10 V to 100 V | $11 \mu\text{V/V} + 0.2 \text{ mV}$ | Fluke 5700A Multi-Function Calibrator | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | 100 V to 1 000 V | $13 \mu\text{V/V} + 0.9 \text{ mV}$ | Fluke 5700A Multi-Function Calibrator | Procedure NS02 | F, O |
| Electrical | Equipment to Output DC Voltage | 1000 V to 10 000 V | 0.35 % of Reading | Fluke 80 Divider | Procedure NS02 | F, O |
| Electrical | Equipment to Measure Resistance | Up to 10 Ω | $20 \mu\Omega/\Omega + 5 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 10 Ω to 100 Ω | $11 \mu\Omega/\Omega + 1 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|--|---|--|--|---|----------------------|
| Electrical | Equipment to Measure Resistance | 100 Ω to 1 k Ω | 11 $\mu\Omega/\Omega + 1 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 1 K Ω to 10 k Ω | 11 $\mu\Omega/\Omega + 1 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 10 K to 100 k Ω | 11 $\mu\Omega/\Omega + 1 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 100 K to 1 M Ω | 13 $\mu\Omega/\Omega + 2 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 1 M to 10 M Ω | 22 $\mu\Omega/\Omega + 12 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 10 M to 100 M Ω | 140 $\mu\Omega/\Omega + 116 \mu\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 100 M Ω to 1 G Ω | 1.5 m $\Omega/\Omega + 1.2 \text{ m}\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Resistance | 1 G Ω to 10 G Ω | 1.5 m $\Omega/\Omega + 1.2 \text{ m}\Omega$ | Fluke 8588 A DMM HP 3458A DMM | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 1 Ω | 133 $\mu\Omega/\Omega + 14 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 1.9 Ω | 131 $\mu\Omega/\Omega + 4 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 10 Ω | 50 $\mu\Omega/\Omega + 4 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 19 Ω | 37 $\mu\Omega/\Omega + 8 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 100 Ω | 24 $\mu\Omega/\Omega + 4 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 190 Ω | 24 $\mu\Omega/\Omega + 4 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 1 k Ω | 18 $\mu\Omega/\Omega + 7 \mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|--|---|--|--|---|----------------------|
| Electrical | Equipment to Output DC Resistance | 1.9 k Ω | 18 $\mu\Omega/\Omega$ + 1 $\mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 10 k Ω | 19 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 19 k Ω | 17 $\mu\Omega/\Omega$ + 5 $\mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 100 k Ω | 19 $\mu\Omega/\Omega$ + 11 $\mu\Omega$ | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 190 k Ω | 19 $\mu\Omega/\Omega$ + 500 m Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 1 M Ω | 27 $\mu\Omega/\Omega$ + 6 Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 1.9 M Ω | 29 $\mu\Omega/\Omega$ + 3 Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 10 M Ω | 55 $\mu\Omega/\Omega$ + 1 k Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 19 M Ω | 69 $\mu\Omega/\Omega$ + 2 k Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Output DC Resistance | 100 M Ω | 161 $\mu\Omega/\Omega$ + 9 k Ω | Fluke 5700A Multi-Function Calibrator | Procedure NS01 | F, O |
| Electrical | Equipment to Measure Current | Up to 100 μ A | 32 μ A/A + 47 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |
| Electrical | Equipment to Measure Current | 100 μ A to 1 mA | 12 μ A/A + 5 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |
| Electrical | Equipment to Measure Current | 1 mA to 10 mA | 11 μ A/A + 5 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |
| Electrical | Equipment to Measure Current | 10 mA to 100 mA | 17 μ A/A + 5 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |
| Electrical | Equipment to Measure Current | 100 mA to 1 A | 66 μ A/A + 12 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|---|---|----------------------|
| Electrical | Equipment to Measure Current | 1A to 10 A | 153 μ A/A + 116 μ A | Fluke 8588 A DMM HP 3458A DMM | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | Up to 100 μ A | 71 μ A/A + 16 nA | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | 100 μ A to 1 mA | 71 μ A/A + 23 nA | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | 1 mA to 10 mA | 54 μ A/A + 162 nA | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | 10 mA to 100 mA | 83 μ A/A + 542 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | 100 mA to 1 A | 112 μ A/A + 92 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current | 1A to 10 A | 6.5 mA | Fluke 5700A Multi-Function Calibrator | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current (Clamp Meters) | 10 A to 200A | 0.14 A | Fluke 5700A Multi-Function Calibrator with Turn Coils | Procedure NS03 | F, O |
| Electrical | Equipment to Output DC Current (Clamp Meters) | 200 A to 700 A | 0.52 A | Fluke 5700A Multi-Function Calibrator with Turn Coils | Procedure NS03 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 10 Hz to 20 kHz) | Up to 100 mV | 48 μ V | HP 3458A DMM Fluke 8588 A DMM | Procedure NS04 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 10 Hz to 20 kHz) | 100 mV to 1 V | 480 μ V | HP 3458A DMM Fluke 8588 A DMM | Procedure NS04 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 10 Hz to 20 kHz) | 1 V to 10 V | 4.8 mV | HP 3458A DMM Fluke 8588 A DMM | Procedure NS04 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 10 Hz to 20 kHz) | 10 V to 100 V | 11 mV | HP 3458A DMM Fluke 8588 A DMM | Procedure NS04 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 45 Hz to 20 kHz) | 100 V to 1000 V | 130 mV | HP 3458A DMM Fluke 8588 A DMM | Procedure NS04 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|--|---|----------------------|
| Electrical | Equipment to Output AC Voltage (@ 30 Hz to 30 kHz) | Up to 100 mV | 43 μ V | Fluke 5700A Multi-Function Calibrator | Procedure NS04 Procedure NS07 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 30 Hz to 30 kHz) | 100 mV to 1 V | 450 μ V | Fluke 5700A Multi-Function Calibrator | Procedure NS04 Procedure NS07 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 30 Hz to 30 kHz) | 1 V to 10 V | 850 μ V | Fluke 5700A Multi-Function Calibrator | Procedure NS04 Procedure NS07 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 30 Hz to 30 kHz) | 10 V to 100 V | 11 mV | Fluke 5700A Multi-Function Calibrator | Procedure NS04 Procedure NS07 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 30 kHz) | 100 V to 1000 V | 55 mV | Fluke 5700A Multi-Function Calibrator | Procedure NS04 Procedure NS07 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 1 MHz) | 3 mV | 0.06 % of Reading + 5 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 10 mV | 0.06 % of Reading + 10 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 KHz to 5 MHz) | 10 mV | 0.06 % of Reading + 37 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 5 MHz to 15 MHz) | 10 mV | 0.06 % of Reading + 79 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 15 MHz to 30 MHz) | 10 mV | 0.06 % of Reading + 0.1 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 30 mV | 0.04 % of Reading + 26 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 100 KHz to 2 MHz) | 30 mV | 0.04 % of Reading + 59 μ V | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 2 MHz to 10 MHz) | 30 mV | 0.04 % of Reading + 0.18 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 MHz to 30 MHz) | 30 mV | 0.052 % of Reading + 0.3 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 100 mV | 0.04 % of Reading + 60 μ V | Fluke 5700 A | Procedure NS04 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|--|---|----------------------|
| Electrical | Equipment to Measure Wideband Flatness (@ 100 KHz to 2 MHz) | 100 mV | 0.2 % of Reading + 0.2 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 2 MHz to 15 MHz) | 100 mV | 0.22 % of Reading + 0.72 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 15 MHz to 30 MHz) | 100 mV | 0.26 % of Reading + 1.2 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 300 mV | 0.03 % of Reading + 0.16 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 100 KHz to 5 MHz) | 300 mV | 0.1 % of Reading + 1 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 5 MHz to 15 MHz) | 300 mV | 0.14 % of Reading + 2.3mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 15 MHz to 30 MHz) | 300 mV | 0.16 % of Reading + 3.6 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 1 V | 0.03 % of Reading + 1 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 100 KHz to 5 MHz) | 1 V | 0.1 % of Reading + 4 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 5 MHz to 30 MHz) | 1 V | 0.2 % of Reading + 13 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 10 Hz to 100 KHz) | 3 V | 0.47 % of Reading + 2 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 100 KHz to 2 MHz) | 3 V | 0.47 % of Reading + 6 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 2 MHz to 15 MHz) | 3 V | 0.48 % of Reading + 22 mV | Fluke 5700 A | Procedure NS04 | F, O |
| Electrical | Equipment to Measure Wideband Flatness (@ 15 MHz to 30 MHz) | 3 V | 0.48 % of Reading + 34 mV | Fluke 5700 A | Procedure NS04 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|---|---|----------------------|
| Electrical | Equipment to Output AC Current (@ 40 Hz to 400 Hz) (Clamp Meters) | 10 A to 200 A | 1.7 A | Fluke 9100 Multi-Function Calibrator with 10 and 50 turns coils | Procedure NS07 | F |
| Electrical | Equipment to Output AC Current (@ 40 Hz to 400 Hz) (Clamp Meters) | 200 A to 700 A | 11 A | Fluke 9100 Multi-Function Calibrator with 10 and 50 turns coils | Procedure NS07 | F |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | Up to 200 μ A | 0.09 % + 0.6 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 100 μ A to 2 mA | 0.09 % + 0.6 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 2 mA to 20 mA | 0.09 % + 6 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 20 mA to 200 mA | 0.09 % + 82 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 200 mA to 2 A | 0.1 % + 0.8 mA | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 10 A | 0.09 % + 0.6 μ A | Fluke 5700A Multi-Function Calibrator | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | Up to 10 μ A | 78 nA | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | 10 μ A to 100 μ A | 520 nA | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | 100 μ A to 1 mA | 7.2 μ A | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | 1 mA to 10 mA | 52 μ A | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | 10 mA to 100 mA | 0.16 mA | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 5 kHz) | 100 mA to 1 A | 7.8 mA | HP3458A DMM Fluke 8588 A DMM | Procedure NS05 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|--|---|----------------------|
| Electrical | Equipment to Measure AC Power (@ 40 Hz to 850 Hz) | 220 V and (up to 7 A) | 0.09 % of Reading + 0.1 W | N4L PPA5510 Power Analyzer | Easy Cal Program | F |
| Electrical | Equipment to Measure AC Power Phase (@ 220 V & 55 Hz) | 0° | 0.2° | N4L PPA5510 Power Analyzer | Easy Cal Program | F |
| Electrical | Equipment to Measure AC Power Phase (@ 220 V & 55 Hz) | 30° | 0.2° | N4L PPA5510 Power Analyzer | Easy Cal Program | F |
| Electrical | Equipment to Measure AC Power Phase (@ 220 V & 55 Hz) | 60° | 0.2° | N4L PPA5510 Power Analyzer | Easy Cal Program | F |
| Electrical | Equipment to Measure AC Power Phase (@ 220 V & 55 Hz) | 90° | 0.2° | N4L PPA5510 Power Analyzer | Easy Cal Program | F |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 1 mV to 10 mV | 50 μ V | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 10 mV to 100 mV | 500 μ V | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 100 mV to 300 mV | 5 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 300 mV to 3 V | 50 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 3 V to 30 V | 300 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (AC Vertical Deflection) | 30 V to 190 V | 750 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Electrical | Oscilloscopes (Bandwidth) (10 MHz to 600 MHz) | 100 mV to 5 V p-p | 0.35 % FS | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |




Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|--|---|--|--|---|----------------------|
| Time and Frequency | Equipment to Generate Frequency (Fixed points) | 1 MHz and 10 MHz | 5 parts in $10^7 + 1$ count | Racal Dana 9475 Rubidium Frequency Standard | Procedure NS06 | F, O |
| Time and Frequency | Equipment to Measure Frequency | 10 Hz to 100 Hz | 10 nHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 100 Hz to 1 KHz | 12 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 1 kHz to 10 KHz | 1.2 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 10 kHz to 100 KHz | 2 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 100 kHz to 1 MHz | 20 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 1 MHz to 5 MHz | 25 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 5 MHz to 10 MHz | 200 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Equipment to Measure Frequency | 10 MHz to 160 MHz | 330 mHz | Agilent 53131A Counter | OEM Manual | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 2 ns to 10 ns | 0.02 ns | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 10 ns to 100 ns | 0.2 ns | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 100 ns to 10 μ s | 0.062 μ s | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 10 μ s to 100 μ s | 20 ns | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 100 μ s to 1 ms | 0.2 ns | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 1 ms to 10 ms | 2 ns | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | LOCATION OF ACTIVITY |
|----------------------|---|---|--|--|---|----------------------|
| Time and Frequency | Oscilloscopes (Time Base) | 10 ms to 100 ms | 20 μ s | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Time Base) | 100 ms to 1 s | 200 μ s | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 1 mV to 10 mV | 50 μ V | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 10 mV to 100 mV | 500 μ V | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 100 mV to 300 mV | 5 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 300 mV to 3 V | 50 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 3 V to 30 V | 300 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (AC Vertical Deflection) | 30 V to 190 V | 750 mV | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |
| Time and Frequency | Oscilloscopes (Bandwidth) (@ 10 MHz to 600 MHz) | (100 mV to 5 V) p-p | 0.35 % FS | Fluke 9500 w/600 MHz Active Head | Procedure NS011 | F, O |



Certificate of Accreditation: Supplement

Repair and Calibration Ltd.

Unit 1 Weighbridge Row, Cardiff Road, Reading, Berkshire, RG1 8LX
Contact Name: Jonathan Lee Phone: 011-895-88391

Accreditation is granted to the facility to perform the following conformity assessment activities:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. Location of activity:

| Location Code | Location |
|---------------|--|
| F | Conformity assessment activity is performed at the CABs fixed facility |
| O | Conformity assessment activity is performed onsite at the CABs customer location |
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.